NOTES:

- 1. SUBSTRATE: FUSED SILICA
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN <3 ARCMIN
- 3. COATING (APPLY ACROSS COATING APERTURE): S1: R(ABS) > 99% @ 355nm T(ABS) > 95% @ 532nm @ 45° AOI

DAMAGE THRESHOLD (PULSED): 2.5 J/cm² @ 20ns , 20 Hz @ 355nm 5 J/cm² @ 20ns , 20 Hz @ 532nm

S2: R(ABS) < 0.25% @ 532nm @ 45° AOI


DAMAGE THRESHOLD (PULSED): 10 J/cm² @ 20ns , 20 Hz @ 532nm




5. ROHS COMPLIANT




ALL DIMS IN



1 OF 1





DWG NO

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2	
SHAPE	PLANO	PLANO	
SURFACE QUALITY	10-5	10-5	
CLEAR APERTURE	Ø21.59	Ø21.59	
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	

			Edmund Optics®	
THIRD ANGLE PROJECTION	ϕ	TITLE	Ø25.4mm HS 355R/532T HARMONIO BEAMSPLITTER	3
ALL DIMS IN	mm	DWG NO	27710	SHEET

37719